



Cascade Engineering Services, Inc.

6640 185th Ave NE, Redmond, WA 98052

T.425.895.8617, F.425.702.9358



ON-SITE CERTIFICATE #: OS3-00324998

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

KITSAP COUNTY SHERIFFS OFFICE

3951 NW RANDALL WAY SILVERDALE, WA 98383

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NC SL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is "As Found" data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCRIPTION:	SPEED MEASURING DEVICE, DOPPLER		
ASSET #:	106	SERIAL NUMBER:	PYT546004101
MANUFACTURER:	MPH INDUSTRIES	MODEL NUMBER:	PYTHON
DEPARTMENT:	N/A	PATROL CAR #:	N/A
ENVIRONMENT:	70.0 °F/50.0 %RH	BASIC ACCURACY:	REFERENCE CAL PROCEDURES
CAL INTERVAL:	24 MONTHS	DUE DATE:	January 22, 2023

EQUIPMENT CONDITION AS RECEIVED

Initial testing found this equipment to be "IN TOLERANCE", as defined by the basic accuracy stated above.

EQUIPMENT CONDITION AS DELIVERED

At the completion of the calibration, measured values were "IN TOLERANCE", as defined by the basic accuracy stated above.

TUNNING FORK(S) SUPPLIED WITH THIS DEVICE

DESCRIPTION	SERIAL NUMBER	RATED SPEED	FREQUENCY
TUNING FORK ONE	070202	35 M.P.H.	2526 Hz
TUNING FORK TWO	070714	65 M.P.H.	4690 Hz

Antenna 1 SN: PYT315011772, Frequency: 24.152 GHz

Antenna 2 SN: PYT315011773, Frequency: 24.154 GHz

STANDARD(S) USED FOR CERTIFICATION

I.D.	MODEL	MANUFACTURER	DESCRIPTION	DUE DATE
MET1231	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFICATION SYSTEM	08/31/2021
MET1232	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND	08/31/2021

PROCEDURE(S) USED FOR CERTIFICATION

DOCUMENT ID	DESCRIPTION	REV	REV DATE
CP-SMD-001	RADAR AND LIDAR CALIBRATION PROCEDURE	NEW	11/14/2019

CERTIFICATION NOTES

I certify (or declare) under penalty of perjury under the laws of the State of Washington that the above information is true and correct

PERFORMED BY 
METROLOGIST: JOHN GRAY

LOCATION: SILVERDALE, WA
CALIBRATION DATE: Friday, January 22, 2021

THIS LABORATORY IS A2LA ACCREDITED TO ISO/IEC 17025:2017 (GENERAL REQUIREMENTS FOR THE COMPETENCE OF TESTING AND CALIBRATION LABORATORIES), CERTIFICATE #: 2560.01